



<http://www.steris-ast.com>

## Certificate of Irradiation

Date Issued: 25-May-2024

MY03S12790144-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products  
 ISO 9001 Quality Management System  
 EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd  
 200 MOO 8 KANCHANAVANICH ROAD  
 PRIK SADAO  
 Amphur Sadad  
 SONGKHLA 90120  
 THAILAND

### Order Information

Account Number:	101195																		
Synergy Health Sales Part Reference:	1126467																		
Customer Reference Number:	4027023921																		
Product Description:	KIMTECH*G3 STERILE LATEX GLOVES,HAND SPECIFIC,12" PAIR PACKED																		
Validation Reference:	0.0764 Rev.02																		
Quantity Received:	216																		
Customer Minimum Specification kGy:	25.0																		
Customer Maximum Specification kGy:	50.0																		
Customer Unit Lot/Batch Number:	SEE BELOW																		
Other Process Details:	Kimtech* G3 Sterile Latex Gloves, Hand Specific,12" Pair Packed KC Code : 56843-40, 56844-40, 56845-40, 56846-40, 56847-40, 56848-40, 56849-40, 56842-40																		
	<table border="1"> <thead> <tr> <th>Catalog Number(s)</th> <th>Lot No./Batch No.</th> <th>Quantity</th> </tr> </thead> <tbody> <tr> <td>56844-40</td> <td>420524/SM41352XX</td> <td>72</td> </tr> <tr> <td>56846-40</td> <td>420524/SM41342XX</td> <td>36</td> </tr> <tr> <td>56846-40</td> <td>420524/SM41272XX</td> <td>24</td> </tr> <tr> <td>56846-40</td> <td>420524/SM41262XX</td> <td>69</td> </tr> <tr> <td>56846-40</td> <td>420524/SM41252XX</td> <td>15</td> </tr> </tbody> </table>	Catalog Number(s)	Lot No./Batch No.	Quantity	56844-40	420524/SM41352XX	72	56846-40	420524/SM41342XX	36	56846-40	420524/SM41272XX	24	56846-40	420524/SM41262XX	69	56846-40	420524/SM41252XX	15
Catalog Number(s)	Lot No./Batch No.	Quantity																	
56844-40	420524/SM41352XX	72																	
56846-40	420524/SM41342XX	36																	
56846-40	420524/SM41272XX	24																	
56846-40	420524/SM41262XX	69																	
56846-40	420524/SM41252XX	15																	

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 1 of 2

*SA. 29/05/2024*  
 NURSYUHADA BT. SHOH  
 Quality Analyst  
 Synergy Sterilisation (M) Sdn. Bhd.  
 +66(0)4415 2127



<http://www.steris-ast.com>

## Certificate of Irradiation

Date Issued: 25-May-2024

MY03S12790144-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products  
 ISO 9001 Quality Management System  
 EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd  
 200 MOO 8 KANCHANAVANICH ROAD  
 PRIK SADAO  
 Amphur Sadad  
 SONGKHLA 90120  
 THAILAND

### Irradiation Data

Date and Time of Irradiation:	24-May-2024 16:28
Reference Dose Range kGy:	34.7 - 35.8
Calculated Minimum Dose kGy:	29.9
Calculated Maximum Dose kGy:	38.3

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 2 of 2

*Suf. 29/05/2024*  
 NURSYUHADA BT. SHOH  
 Quality Analyst  
 Synergy Sterilisation (M) Sdn. Bhd.  
 +60(0)4415 2127